

## Declaration of Conformity

### The Directives covered by this declaration:

2014/30/EU Electromagnetic Compatibility

Declares that the product(s): **Elcometer NDT Ultrasonic Bolt Tension Monitors**

Part Number(s): **BG80DL, BG80TDL**

Product Option(s): **TL-24030-3, TL-24030-(10-12), TL-24031, TL-24032, TX10M0AM-(2-3), TX10M0CM-(4-5), TX2M25CM-(5-6), TX2M25DM, TX2M25EM-(5-6), TX2M25GM, TX5M00AM-(2-3), TX5M00CM-(7-8), TX5M00DM, TX5M00EM-(6-7), TX5M00GM, TZ-24039**

This declaration of conformity is issued under the sole responsibility of Elcometer Limited. The product identified above complies with the requirements of the above EU Directive by meeting the following standards:

<b>EN 61326-1:2013<sup>1</sup></b> <b>IEC 61326-1:2012</b> <b>Class A<sup>2</sup>, Group 1<sup>3</sup> ISM</b>	Electrical equipment for measurement, control and laboratory use – EMC requirements Part 1 General requirements.
--	---

<sup>1</sup> CKC Laboratories Inc Test Report 92753-7

<sup>2</sup> Class A product: Suitable for use in all establishments other than domestic and those directly connected to a low voltage power supply network which supplies buildings used for domestic purposes.

<sup>3</sup> Group 1 ISM product: Product in which there is intentionally generated and/or used conductively coupled radio-frequency energy which is necessary for the internal functioning of the equipment itself.



Signed:

M. C. Sellars

Manchester, UK

Date: 4<sup>th</sup> July 2016

Authority: Managing Director