

Elcometer Limited Edge Lane, Manchester M43 6BU. England Tel: +44 (0)161 371 6000 Fax: +44 (0)161 371 6010 sales@elcometer.com www.elcometer.com REGISTERED NO. 1729726 ENGLAND Accredited to ISO 9001 ISO 14001

## Declaration of Conformity

## The Directives covered by this declaration:

2014/30/EU Electromagnetic Compatibility

## Declares that the product(s): Elcometer NDT FD700DL+ Flaw Detection Gauge

Part Number(s): FD700DL+

Product Option(s): TF1M00XXX-(1 or 2), TF2M25XXX-(1 or 2), TF3M50XXX-(1 or 2), TF5M00XXX-(1 or 2), TF15M0XXX-(1 or 2), TF20M0XXX-(1 or 2), TF9999FF45-1, TF9999FF60-1, TF9999FF70-1, TL-24030-(1-3), TL-24030-(5-8), TL-24031, TL-24032, TX5M00GM

This declaration of conformity is issued under the sole responsibility of Elcometer Limited. The product identified above complies with the requirements of the above EU Directive by meeting the following standards:

EN 61326-1:2013 <sup>1</sup>	Electrical equipment for measurement, control
IEC 61326-1:2012	and laboratory use – EMC requirements
Class A <sup>2</sup> , Group 1 <sup>3</sup> ISM	Part 1 General requirements.

<sup>&</sup>lt;sup>1</sup> CKC Laboratories Inc Test Report 92005-4

Signed:

M. C. Sellars

Manchester, UK Date: 4<sup>th</sup> July 2016 Authority: Managing Director

<sup>&</sup>lt;sup>2</sup> Class A product : Suitable for use in all establishments other than domestic and those directly connected to a low voltage power supply network which supplies buildings used for domestic purposes.

<sup>&</sup>lt;sup>3</sup> Group 1 ISM product: Product in which there is intentionally generated and/or used conductively coupled radio-frequency energy which is necessary for the internal functioning of the equipment itself.